



**First Quality  
in Ophthalmology**

1stQ GmbH | Harrlachweg 1 | 68163 Mannheim

1stQ GmbH  
Harrlachweg 1  
68163 Mannheim  
T. +49 (0) 621 7895-3790  
F. +49 (0) 621 7895-3791

22. Februar 2017

## **Field Safety Notice**

### **Voluntary field Safety Corrective Action:**

#### **Pseudophakic intraocular lens A4SW00 Acrylic IOL, AddOn, Spherical, hydrophilic, clear; -10,0 D to +10,0 D**

Dear Sir or Madam,

With this letter we inform you about the voluntary recall of spherical AddOn intraocular lenses (IOL), article number: A4SW00.

**1stQ has identified involved articles in your organization as listed in attachment.**

#### **Description of the problem and the identified cause:**

1stQ conducts this recall, after a few single cases of fibrin reactions, following implantation of spherical AddOn lenses, have been reported.

These single cases, correlated to the entire amount of spherical AddOn IOL implanted, do not necessarily indicate that the AddOn IOL is the cause of the fibrin reaction.

Due to the symptoms described, TASS (toxic anterior segment syndrome) is likely to be the reason for the fibrin reaction. TASS is described as an inflammatory reaction restricted to the anterior chamber of the eye, which is caused by noninfectious foreign particles within the eye<sup>1,2</sup>. The cases described in the literature identify different products for cataract surgery as reasons for TASS. Therefore, it is currently difficult to specify the definite cause.

In order to exclude any risk for patients, all spherical AddOn IOLs listed in table 1 are recalled. The recall applies only to spherical AddOn IOLs, which have been produced from a particular lot of hydrophilic acrylate.

Toric and progressive AddOn intraocular lenses are not affected by the recall.

#### **Required Measures**

Seite 2

22. Februar 2017

Block all lenses with the mentioned serial numbers. In case one of those lenses has been implanted already, it is recommended, to observe the patient for one month post-operatively. If no symptoms are observed, no further measures have to be taken.

If any TASS cases occur, please report this either to us or to the responsible health authority.

Please return all acrylic IOLs, AddOn, spherical, hydrophilic, clear, A4SW00, possessing a serial number listed in the attachment and the completed response form to the following address:

**Contact data:**

1stQ GmbH  
Dr. M. Kirchenbauer  
Harrlachweg 1  
68163 Mannheim  
Germany

You may return the completed response form via Fax: 0049 (0)621 71763-33 or Email:  
[kirchenbauer@1stq.de](mailto:kirchenbauer@1stq.de).

Replacement or credit will take place after consultation.

1stQ GmbH apologizes for any inconveniences that may occur in relation to this measure. 1stQ GmbH places emphasis on high quality standards, consequently, this measure will be conducted with utmost care.

**Distribution of relevant information:**

Please ensure, that all users and relevant persons in your organization are informed about this field safety notice. In case you distributed this product to a third party, please forward a copy of this FSN or inform the contact person indicated below.

**Notification to authorities**

1stQ GmbH has informed the responsible national authority.

Kind regards

Dr. M. Kirchenbauer

<sup>1</sup> Lucien A. M. van Philips; Toxic Anterior Segment Syndrome after Foldable Artiflex Iris-Fixated Phakic Intraocular Lens Implantation; Journal of Ophthalmology; Volume 2011, Article ID 982410, 5 pages

<sup>2</sup> Kremer I et al; Toxic anterior segment syndrome following iris-supported phakic IOL implantation with viscoelastic Multivisc BD.; [Eur J Ophthalmol](#). 2010 Mar-Apr;20(2):451-3.

**Please complete the response form with listed items in the attachment.**

**Name and title of the person completing the form:**

**Organisation:**

Please return the completed form via Fax: 0049 (0)621 71763-33 or Email:  
[kirchenbauer@1stq.de](mailto:kirchenbauer@1stq.de)