RAND®X
Urgent Field Safety Notice

Date: 25th April 2017

Complaint Reference: 287 **Action Type:** Device Modification

Detail on Affected Devices:

Our records indicate that your facility may have received the following product.

Assay Catalogue GTIN Batch/Lot Expiry Date Manufacturing Date

Number CKMB CQ5053 5055273207460 3912CK 28th October 2017 7th November 2016

Reason for Recall:

Randox has confirmed a change in recovery with regards to CKMB in lot 3912CK of the Liquid Cardiac Control. A similar issue with Myoglobin has previously been reported for this lot. All other analytes recover as expected.

The IFU has been updated to indicate in the limitations section that values for CKMB and Myoglobin may gradually decrease over the product shelf-life for the Liquid Cardiac Controls (CQ505 1, CQ5052 and CQ5053).

Risk to Health:

The quality control results which are not within range can lead to delay in reporting CKMB results however as CKMB is no longer used as the primary test for diagnosing cardiac injury this therefore should not pose a serious risk to health.

Action to be taken:

• Place a copy of the important notice in all kits.

• Discuss the contents of this notice with your Medica! Director.

• Compliance with your country's Regulatory Authority requires a return of the attached response form. Please complete and return the vigilance response section of this form to technical.services@randox.com within five working days.

RAND®X Urgent Field Safety Notice

Transmission of Field Safety Notice: Send a copy of the FSN to all affected customers and to those who need to be aware within your organisation.

Contact Reference:

Randox Technica! Services Randox Laboratories Ltd, 55 Diamond Road, Crumlin, United Kingdom, BT294QY

Email: techn ical.services@randox.com

Tel: +44 (0) 28 9445 I 070 Fax: +44 (0) 28 9445 2912

.

Please accept our apologies for any inconvenience caused. Thank you for your patience and understanding. Ifyou have any questions or concerns please contact Randox Technica! Services.

The undersigned confirms that this notice has been notified to the appropriate Regulatory Agency

April 2017

RAND®X Urgent Field Safety Notice

Vigilance Response	Form (Response Plan	n must be complete	ed by the importer of	f the device)
Importer Details				
Company Name				
Address				
Total Quantity				
Received				
Distributed				
Area of Distribution				
(To be completed by	Distributors and Ran	dox Offices)		
Consignee	Country	Quantity Received	Analyser Serial Number	Replacements Required
I have read and under	rstood the Urgent Fie	ld Safety Notice. T	The actions to be take	en are completed.
Completed By			Date +	
	Tel	Email	. 1	