



URGENT MEDICAL DEVICE CORRECTION

GE Healthcare

Healthcare Digital
500 W. Monroe St.
Chicago, IL 60661 USA

<Date of Letter Deployment>

GEHC Ref# 85443

To: Director/Manager of Radiology
Hospital Administrator
Head of Radiology Department
PACS Administrator
Director of IT Department

RE: Centricity PACS-IW versions 3.6.0 through 3.7.3.9 SP1, SP2, SP3 and 3.7.3 SPa10 potential that one or more “image series” (i.e. all images within an image set) may be missing from an exam.

Please ensure that all potential users in your facility are made aware of this safety notification and the recommended actions.

Safety Issue

A database handling error could occur during the image acquisition process affecting the completeness of acquired images with Centricity PACS-IW. There is a potential that one or more “image series” (i.e. all images within an image set) may be missing from an exam without indication to the user. While this is rare, this can occur with imaging studies that consist of a very small number of images per series. E.g. CR Thorax exam with 1 image per series. There have been no injuries reported as a result of this issue.

Safety Instructions

Until a product modification is available to correct this issue, the following actions are recommended:

1. Utilize the image count within the QC process to alert the user of a discrepancy in the number of transmitted images from the modality to the number of images available in the Universal Viewer.
 - a. If a discrepancy is identified, attempt to retransmit the exam to PACS.
 - b. If retransmission is unsuccessful, a GE Healthcare Service representative will need to be contacted to assist with resolution of the impacted exam. Urgent cases impacted by this issue should be interpreted at the modality.
2. If DICOM storage commit is configured and in use, no commit notification is sent to the modality for images that were impacted by this issue.
3. Attention should be given to prior cases, acquired before the application of the safety instructions given in this communication as they may be affected by the safety issue.

Affected Product Details

Centricity PACS-IW versions 3.6.0 through 3.7.3.9 SP1, SP2 and SP3, 3.7.3 SPa10.

Product Correction

GE Healthcare will correct all affected systems by providing a product correction at no cost to you. A GE Healthcare representative will contact you to arrange for the correction.

Contact Information

If you have any questions or concerns regarding this notification, please contact GE Healthcare Service at 1-800-437-1171 or your local Service Representative.

Please be assured that maintaining a high level of safety and quality is our highest priority. If you have any questions, please contact us immediately per the contact information above.

Sincerely,

....
....
GE Healthcare

....
....
GE Healthcare