

FSN Ref: WW20_C11607-DP



FSCA Ref: C11607-DP


Date: 11.08.2020

Urgent Field Safety Notice
Dispenser DP 30 LipoPlus

For Attention of: Emdaplast – Netherlands

Contact details of local representative
Nouvag AG [REDACTED] St. Gallerstrasse 23-25 9403 Goldach +41 71 846 66 57



Urgent Field Safety Notice (FSN) DP 30 LipoPlus
Production according to expired EMV Standard 60601-1-2 Edition 3

1. Information on Affected Devices	
1.	<p>1. Device Type</p> <p>The Dispenser DP 30 LipoPlus is a specifically for liposuction designed tumescence infiltration pump, delivering high volume of tumescence liquid.</p> <div style="text-align: center;">  </div>
1.	<p>2. Commercial name(s)</p> <p>Dispenser DP 30 LipoPlus</p>
1.	<p>3. Unique Device Identifier(s) (UDI-DI)</p> <p>+ENOU41610F +ENOU41630H</p>
1.	<p>4. Primary clinical purpose of device(s)*</p> <p>The DP 30 LipoPlus is a mobile Infiltration pump that is used for Tumescence infiltration during Liposuction and for treatments in Angiology</p>
1.	<p>5. Device Model/Catalogue/part number(s)</p> <p>4161 and 4163</p>

1.	6. Affected serial or lot number range		
	Qty	SET SN	UNIT SN
	1	6638S1905R	7911U1901R
	1	2391S1909R	4620U1904R

2 Reason for Field Safety Corrective Action (FSCA)	
2.	1. Description of the product problem The devices DP 30 LipoPlus does not comply with the latest harmonized EMC standard (60601-1-2, Edition 4). The device only complies with the expired Edition 3 and was not adapted to the new standard.
2.	2. Hazard giving rise to the FSCA The device might interfere with other electrical devices. The DP 30 LipoPlus could disturb the function of devices nearby or could itself be disturbed by them.
2.	3. Probability of problem arising Little to no probability of problems arising. The device still complies with the previous Edition 3 EMC standard (IEC 60601-1-2:2007). With the harmonization of the EMC standard Edition 4 (IEC 60601-1-2:2014) the acceptable ranges of electromagnetic interference is now smaller and thus not successfully achieved by the device.
2.	4. Predicted risk to patient/users none

3. Type of Action to mitigate the risk			
3.	<p>1. Action To Be Taken by the User</p> <p> <input checked="" type="checkbox"/> Identify Device <input type="checkbox"/> Quarantine Device <input checked="" type="checkbox"/> Return Device <input type="checkbox"/> Destroy Device </p> <p> <input type="checkbox"/> On-site device modification/inspection </p> <p> <input type="checkbox"/> Follow patient management recommendations </p> <p> <input type="checkbox"/> Take note of amendment/reinforcement of Instructions For Use (IFU) </p> <p> <input type="checkbox"/> Other <input type="checkbox"/> None </p>		
	<p>Device must be returned to the following address:</p> <p> Nouvag GmbH Dental und Medizintechnik Schulhausstrasse 15 DE - 78462 Konstanz Germany </p> <p> Tel. +49 (0)7531 1290-0 Fax +49 (0)7531 1290-12 info-de@nouvag.com </p>		
3.	<table border="1" style="width: 100%;"> <tr> <td style="width: 40%;">2. By when should the action be completed?</td> <td>Immediately</td> </tr> </table>	2. By when should the action be completed?	Immediately
2. By when should the action be completed?	Immediately		
3.	<table border="1" style="width: 100%;"> <tr> <td style="width: 60%;">3. Is customer Reply Required? * (If yes, form attached specifying deadline for return)</td> <td>Yes, As soon as possible</td> </tr> </table>	3. Is customer Reply Required? * (If yes, form attached specifying deadline for return)	Yes, As soon as possible
3. Is customer Reply Required? * (If yes, form attached specifying deadline for return)	Yes, As soon as possible		
3.	<p>4. Action Being Taken by the Manufacturer</p> <p> <input type="checkbox"/> Product Removal <input type="checkbox"/> On-site device modification/inspection <input type="checkbox"/> Software upgrade <input type="checkbox"/> IFU or labelling change <input checked="" type="checkbox"/> Other <input type="checkbox"/> None </p> <p>Device modification on manufacturing site</p>		

4. General Information	
4.	1. FSN Type New
4.	2. For updated FSN, reference number and date of previous FSN N/A
4.	3. For Updated FSN, key new information as follows: N/A
4.	4. Manufacturer information (For contact details of local representative refer to page 1 of this FSN)
	a. Company Name Nouvag AG
	b. Address St. Gallerstrasse 23-25, CH-9403 Goldach
	c. Website address www.nouvag.com
4.	5. The Competent (Regulatory) Authority of your country has been informed about this communication to customers.
4.	6. Name/Signature 
	

Transmission of this Field Safety Notice	
	<p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)</p> <p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback.</p>

It is important that your organisation takes the actions detailed in the FSN and confirms that you have received the FSN.

Your organisation's reply is the evidence we need to monitor the progress of the corrective actions.

Please fill in the customer/ distributor reply form and send it to us before the defined deadline at: vigilance@novag.com