



1.10.2020

## URGENT FIELD SAFETY NOTICE

Product name: Actim CRP

Date: 1<sup>st</sup> October 2020

Product name (catalogue number)	Lot numbers
Actim CRP (31031ETAC)	0044318, 0044623, 0044821

### Dear Receiver,

The purpose of this letter is to inform you of a recall for product correction for the above products.

### Description of the problem

We have observed that in **occasional Actim CRP dipsticks**, the third test line (>80 mg/L) is lighter than specified. This means that the CRP concentrations >80 mg/L can be interpreted as concentrations between 40 and 80 mg/L. This CRP level is already an indication for further testing or treatment and there is not considered to be a risk of health incidence and no extra follow up of patients is needed. To ensure that no false patient results are observed, **we have decided to withdraw the kit lots affected by this problem** (see above).

### Actions required from receiver

1. Confirm via email that you have received this information.
2. Inform all your customers of this information of withdrawal.
3. Advise your customers to discard the kits that have been withdrawn due to this notification.
4. Assess the number of kits delivered to your customer. Please fill information of all your lots to the "Distributor verification form"
5. Assess the number of kits in your storage. Please fill information of all your lots to the "Distributor verification form"
6. Send sample kit of some of the affected lots for our investigation (if available).
7. Complete "Distributor verification form" and email to Actim [actim@actimtest.com](mailto:actim@actimtest.com) latest 15<sup>th</sup> Oct 2020.

Actim – a part of Medix Biochemica

Headquarters: Klovipellontie 3, FI-02180 Espoo, Finland

Manufacturing site: Nolljakantie 13, FI-80130 Joensuu, Finland

[actim@actimtest.com](mailto:actim@actimtest.com)

[www.actimtest.com](http://www.actimtest.com)

VAT reg.no. FI29540422



1.10.2020

**Transmission of this Field Safety Notice**

This notice needs to be passed on all those who need to be aware within your organization or to any organization where the potentially affected devices have been transferred.

Please transfer this notice to other organizations on which this action has an impact.

Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.

The undersign confirms that this notice has been notified the appropriate National Competent Authorities.

Please accept our sincere apology for all the inconvenience this unfortunate situation brings to you.

If you have any questions or concerns, please do let us know.

[Redacted signature]

Date and signature

**Contact reference person:**

[Redacted contact information]

Actim Oy

Klovinpellontie 3, FI-02180 Espoo, Finland

Tel. +358 9 457 68172

Mobile [Redacted]

Fax +358 9 505 3441

Email: [Redacted]

[www.actimtest.com](http://www.actimtest.com)

**Actim – a part of Medix Biochemica**

Headquarters: Klovinpellontie 3, FI-02180 Espoo, Finland

Manufacturing site: Noljakantie 13, FI-80130 Joensuu, Finland

[actim@actimtest.com](mailto:actim@actimtest.com)  
[www.actimtest.com](http://www.actimtest.com)

VAT reg.no. FI29540422