

Rev 1: September 2018

FSN Ref: FSN-2021-010 FSCA Ref: FSN-2021-010

Date: 5 October 2021

Urgent Field Safety Notice
Pathodxtra Strep Group D Latex

For Attention of*: Lab Managers

Contact details of local representative (name, e-mail, telephone, address etc.)*
--

E.mail : mbd.vigilance@thermofisher.com

Telephone: +44(0) 1256 841144

Fax: +44(0) 1256 479525

Urgent Field Safety Notice (FSN)
Pathodxtra Strep Group D Latex

1. Information on Affected Devices*	
1.	1. Device Type(s)* IVD
1.	2. Commercial name(s) Pathodxtra Strep Group D Latex
1.	3. Unique Device Identifier(s) (UDI-DI) 05032384519866
1.	4. Primary clinical purpose of device(s)* PathoDxtra™ supplementary Strep Grouping reagents are intended for use in the PathoDxtra Strep Grouping Kits, which contain all components necessary for the various systems. These reagents are provided as replacement items. For complete information and correct use of this product, please refer to the Instructions for Use which accompanies the PathoDxtra kits (Strep Grouping Kits [DR0700M and DR0710M]).
1.	5. Device Model/Catalogue/part number(s)* DR0704G
1.	6. Software version N/A
1.	7. Affected serial or lot number range 3312966 & 3312967
1.	8. Associated devices N/A

2. Reason for Field Safety Corrective Action (FSCA)*	
2.	1. Description of the product problem* An internal technical investigation has determined that DR0704G Pathodxtra Strep Group D Latex Lots. 3312966 and 3312967 may fail to agglutinate with Sero type D and therefore lead to false negative results.
2.	2. Hazard giving rise to the FSCA* False Negatives
2.	3. Probability of problem arising High
2.	4. Predicted risk to patient/users There should be no immediate or long-term health consequences from use of this Group D streptococcal testing product. This was found from an internal investigation. Primary quality control would identify failure to agglutinate with the standard quality control strains of enterococci, thus preventing use on clinical isolates. Other standard tests as well (e.g. gram smear, bile esculin etc.) would identify the failure of the defective PathoDx Group D reagent. The clinical risk should therefore be considered as negligible.
2.	5. Further information to help characterise the problem N/A
2.	6. Background on Issue

Rev 1: September 2018

FSN Ref: FSN-2021-010 FSCA Ref: FSN-2021-010

4. General Information*		
4.	1. FSN Type*	New
4.	2. For updated FSN, reference number and date of previous FSN	N/A
4.	3. For Updated FSN, key new information as follows:	
	N/A	
4.	4. Further advice or information already expected in follow-up FSN? *	Not planned yet
4	5. If follow-up FSN expected, what is the further advice expected to relate to:	
	N/A	
4	6. Anticipated timescale for follow-up FSN	N/A
4.	7. Manufacturer information (For contact details of local representative refer to page 1 of this FSN)	
	a. Company Name	Thermo Fisher Scientific
	b. Address	Clipper Boulevard West, Cross ways industrial estate, Dartford, Kent. DA2 6PT
	c. Website address	www.thermofisher.com
4.	8. The Competent (Regulatory) Authority of your country has been informed about this communication to customers. *	
4.	9. List of attachments/appendices:	Customer response form
4.	10. Name	... Vice President, Quality and Regulatory, MBD
	Signature	...

Transmission of this Field Safety Notice	
	<p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)</p> <p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback.*</p>

Rev 1: September 2018

FSN Ref: FSN-2021-010 FSCA Ref: FSN-2021-010

Customer Reply Form

1. Field Safety Notice (FSN) information			
FSN Reference number*	FSN-2021-010		
FSN Date*	05 October 2021		
Product/ Device name*	Pathodxtra Strep Group D Latex		
Product Code(s)	DR0704G		
Batch/Serial Number (s)	3312966 & 3312967		
2. Customer Details			
Account Number			
Organisation Name*			
Organisation Address*			
Department/Unit			
Shipping address if different to above			
Contact Name*			
Title or Function			
Telephone number*			
Email*			
3. Customer action undertaken on behalf of Healthcare Organisation			
<input type="checkbox"/>	I confirm receipt of the Field Safety Notice and that I read and understood its content.		
<input type="checkbox"/>	I performed all actions requested by the FSN.		
<input type="checkbox"/>	The information and required actions have been brought to the attention of all relevant users and executed.		
<input type="checkbox"/>	I have returned affected devices - enter number of devices returned and date complete or N/A	Qty:	Lot/Serial Number: Date Returned (DD/MM/YY)
		Comments:	
<input type="checkbox"/>	I have destroyed affected devices – enter number destroyed and date complete. (EDIT WHEN NECESSARY) or N/A	Qty:	Lot/Serial Number: Date Returned (DD/MM/YY)
		Qty	Credit <input type="checkbox"/> Replacement <input type="checkbox"/>
		Comments:	
<input type="checkbox"/>	No affected devices are available for return/ destruction		
<input type="checkbox"/>	Other Action (Define):		
<input type="checkbox"/>	I do not have any affected devices.		
<input type="checkbox"/>	I have a query please contact me (e.g. need for replacement of the product).		
Print Name*			
Signature*			
Date*			
4. Return acknowledgement to sender			
Email	MBD.vigilance@thermofisher.com		
Telephone Number & Fax	Tel : +44(0) 1256 841144 Fax :+44(0) 1256 479525		
Postal Address			
Deadline for returning the reply form*	02 November 2021		

Mandatory fields are marked with *

It is important that your organisation takes the actions detailed in the FSN and confirms that you have received the FSN. Your organisation's reply is the evidence we need to monitor the progress of the corrective actions.