

ANGIOKARD Medizintechnik GmbH [REDACTED]

**Safety notice - Recall**[REDACTED]  
[REDACTED]  
[REDACTED]

Friedeburg, 26.02.2024

**Urgent safety notice (product recall)**

Dear,

today we have to inform you about the following precautionary product recall:

<b>affected products</b>	Angiokard - articles containing or fitted with a 1-way stopcock which cannot be reliably connected via an LL connection despite the screw cap REF. no. see appendix 1
<b>batch(es) concerned</b>	see appendix 1
<b>AK reference no.</b>	2024-01
<b>cause of FSN</b>	male side of the 1-way valve cannot be securely connected
<b>explanation</b>	via a customer complaint has made us aware that the 1-way valve cannot be reliably and securely connected to other components via the LL connections
<b>expected risk / damage</b>	delay or termination of operations

**This results in the following measures for you:**

- no longer use the affected product(s)
- return products from your stock
- inform all affected persons (users, customers, etc.)
- fill out the confirmation form and return it

We thank you in advance for your cooperation and ask for your understanding.

With kind regards

Angiokard Medizintechnik GmbH

i.V.

[REDACTED]

**Attachment:**

- 1) confirmation form
- 2) list of REF numbers and batches concerned

[REDACTED]

**Urgent safety notice (product recall)**

<b>affected products:</b>	Angiokard – sets with affected 1-way stopcock
<b>cause for FSN:</b>	male side of the 1-way stopcock is not safe to connect

**(by E-Mail with signed PDF to [recall@be.lrmed.com](mailto:recall@be.lrmed.com) by 31.03.2024)**

\* No refund is possible for later responses

**sender:**

**Angiokard Medizintechnik GmbH**

**recipient:**

& all users who use the named products

**Corrective action: Please stop using the affected products.**

For confirmation and for your refund, please provide the following information or acknowledgement as a customer, distributor or retailer of the affected product batch(es).

I hereby confirm, that the affected products are no longer in use. In addition, all persons involved (including third parties, if they have been supplied with the affected products) have been informed of this urgent safety notice.

*Please check one option:*

- the affected products have been used up.
- the remaining products should be returned. After return, the returned goods will be credited. [including those that have been passed on to third parties]

Date / Signature: \_\_\_\_\_

Name in Block capitals: \_\_\_\_\_

Position: \_\_\_\_\_

Department / Institution: \_\_\_\_\_

Phone and Email: \_\_\_\_\_

Please enter the returned products with quantities in the following table:

Appendix 2 Individual customer list with affected REFs and batches

REF (SAP-Nr.)	produkt	batche(s)	quantity Return (pcs.)
161156	Angiographieset Radiologie	239181-000 238869-000	

