



PERI-GUARD Repair Patch and SUPPLE PERI-GUARD Repair Patch

FA-2024-010

Manufacturer: SYNOVIS LIFE TECHNOLOGIES INC. (ST. PAUL) (SRN US-MF-000028264)

Safety Alert

February XX, 2024 (to be adapted locally)

Dear Sir/Madam (to be adapted locally),

**Problem
Description**

On March 7, 2022, Baxter implemented labeling changes on CE-marked **Peri-Guard Repair Patch** and **Supple Peri-Guard Repair Patch** products including removal of the indications for abdominal wall defect and hernia (diaphragmatic, femoral, incisional, inguinal, lumbar, paracolostomy, scrotal and umbilical) repair from the Instruction for Use (IFU).

The indications on the labeling changed from:

For use as a prosthesis for pericardial closure and soft tissue deficiencies which include: defects of the abdominal and thoracic wall, hernias (diaphragmatic, femoral, incisional, inguinal, lumbar, paracolostomy, scrotal and umbilical), and intracardiac and great vessel repair.

To:

For use as a prosthesis for pericardial closure and soft tissue deficiencies which include: defects of the thoracic wall, and intracardiac and great vessel repair.

Baxter has since received seven (7) complaints reporting infection/abscess, all of which were received from one and the same clinic in Italy. Six (6) of the complaints were determined to be correlated to “off-label” use of **Peri-Guard Repair Patch** (multiple product codes) in abdominal surgery.

As these changes in indications may not have been taken into account by all product users, Baxter is informing customers of the changed intended use (limitation of the use) to ensure the correct use of the devices in the market.

**Affected Product
(to be adapted
locally)**

Product Code	Description	Lot Number
PC0404N	Peri-Guard Repair Patch, 4x4cm	All within expiry
PC0608N	Peri-Guard Repair Patch, 6x8cm	All within expiry
PC0814N	Peri-Guard Repair Patch, 8x14cm	All within expiry



PC1016N	Peri-Guard Repair Patch, 10x16cm	All within expiry
PC1225N	Peri-Guard Repair Patch, 12x25cm	All within expiry
PC0404SN	Supple Peri-Guard Repair Patch, 4x4cm	All within expiry
PC0608SN	Supple Peri-Guard Repair Patch, 6x8cm	All within expiry
PC0814SN	Supple Peri-Guard Repair Patch, 8x14cm	All within expiry
PC1016SN	Supple Peri-Guard Repair Patch, 10x16cm	All within expiry

Hazard Involved The change was not driven by any known safety concerns. However, there is a lack of clinical data supporting the safety and effectiveness of the Peri-Guard Repair Patch and Supple Peri-Guard Repair Patch for abdominal wall and hernia defect repair. As such, these indications are no longer approved for CE-Marked Peri-Guard Repair Patch and Supple Peri-Guard Repair Patch and should be considered “off-label” in the European Union.

Action to be taken by the user

Baxter is kindly asking that you take the following actions:

1. Clinicians may continue to use the Peri-Guard Repair Patch and Supple Peri-Guard Repair Patch products listed above however, clinicians should be aware of the recent removal of the product abdominal wall defect and hernia repair indications from the IFUs.
2. Complete the enclosed customer reply form and return it to Baxter by either faxing it to (insert local contact information) or scanning and e-mailing it to (insert local contact information) or sending it by post to (insert local contact information), even if you don't have any inventory. Returning the customer reply form promptly will confirm your receipt of this notification and prevent you from receiving repeat notices.
3. If you purchased this product from a distributor, please note that the Baxter customer reply form is not applicable. If a reply form is provided by your distributor or wholesaler, please return it to the supplier according to their instructions.



4. If you distribute this product to other facilities or departments within your institution, please forward a copy of this communication to them.
5. If you are a dealer, wholesaler, distributor/reseller, or original equipment manufacturer (OEM) that distributed any affected product to other facilities, please notify your customers of this *Safety Alert* in accordance with your customary procedures.

Further information and support (to be adapted locally)

For general questions regarding this communication or any product issue you are experiencing, contact Baxter at (insert local contact information), between the hours of (insert local information).

The local Ministry of Health (MOH) has been notified of this action. (to be adapted locally)

We appreciate your attention to this matter and apologize for any inconvenience this may cause you and your staff.

Sincerely,

Name (to be adapted locally)

Title (to be adapted locally)

Baxter Healthcare Corporation (to be adapted locally)