

Serumwerk Bernburg AG - [redacted]

To:
Pharmacies, distributors, hospitals, that
have purchased EmboLog S 450 mg/
7,5 ml, #00124/#00222

Batch recall notification for EmboLog S 450 mg/ 7.5 ml Amilomer, DSM 35/50

Article no. 3324, Ch.-B.: 00124/00222
Complaint number: 2025-18
PZN - 15885398

Dear Sir or Madam,

In the period March 2024/May 2025, you ordered our product

EmboLog S 450 mg/ 7.5 ml Amilomer, DSM 35/50 article no. 3324, batch numbers 00124,
00222 expiry date 11-30-2025 from Serumwerk Bernburg AG.

Due to a complaint, it was determined that isolated bottles of the above batch may contain
small black particles. It cannot be ruled out that other units may also be affected.

For reasons of medical product safety, we are recalling the above batch on our own
responsibility and as a precautionary measure.

Please check your stocks immediately and inform us accordingly.
We also ask you to inform your customers and, if necessary, to recall the product.
We will then collect the stock from you.

We apologize for any inconvenience caused.

With kind regards

Serumwerk Bernburg AG

Please confirm receipt by June 30, 2025 by fax to [redacted]

or by email: [redacted]

We have properly received the notification.

Date: _____

Time: _____

Stamp, signature